
Certificate of Analysis

Product Information

Part Number : 910107
 Description : PH7.00 BUFFER COLOR - YELLOW, 475ML
 Lot Number : SBS6
 Expiration Date : 21-Aug-2025
 Method of Analysis : POTENTIONMETRIC

Actual Analysis

Acceptance Range	Results	Unit of Measure
7.00 +/- 0.01 PH UNITS @ 25°C	7.00	PH @ 25°C

Traceability

N.I.S.T. Chemical	SRM#
POTASSIUM DIHYDROGEN PHOSPHATE	186-I-g
DISODIUM HYDROGEN PHOSPHATE	186-II-g
SODIUM BICARBONATE	191d-I
SODIUM CARBONATE	191d-II

Certified By : Thermo Fisher Scientific
 Water and Lab Products
 Quality Assurance

Certificate Date : 21-Aug-2023

Find out more at [thermofisher.com/water](https://www.thermofisher.com/water)

Contact us today: wlp.techsupport@thermofisher.com