
Certificate of Analysis

Product Information

Part Number : 910110
Description : PH10.01 BUFFER COLOR - BLUE, 475ML
Lot Number : SBS6
Expiration Date : 22-Aug-2025
Method of Analysis : POTENTIONMETRIC

Actual Analysis

Acceptance Range	Results	Unit of Measure
10.01 +/- 0.02 PH UNITS @ 25°C	10.01	PH @ 25°C

Traceability

N.I.S.T. Chemical	SRM#
POTASSIUM DIHYDROGEN PHOSPHATE	186-I-g
DISODIUM HYDROGEN PHOSPHATE	186-II-g
SODIUM BICARBONATE	191d-I
SODIUM CARBONATE	191d-II

Certified By : Thermo Fisher Scientific
Water and Lab Products
Quality Assurance

Certificate Date : 22-Aug-2023

Find out more at [thermofisher.com/water](https://www.thermofisher.com/water)

Contact us today: wlp.techsupport@thermofisher.com

For Research Use Only. Not for use in diagnostic procedures. © 2020 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific and its subsidiaries unless otherwise specified.