
Certificate Of Analysis

Product Information

Part Number: 910460
Description: PH 4.01 BUFFER, COLOR-RED, 5X60ML
Lot Number: BT2B
Expiration Date: 07/31/2025
Method Of Analysis: POTENTIOMETRIC

Actual Analysis

Acceptance Range	Results	Unit Of Measure
4.01 +/- 0.01 PH UNITS AT 25 C	4.00	PH UNITS AT 25 C

Traceability

N.I.S.T Chemical	SRM#
DISODIUM HYDROGEN PHOSPHATE	186 II g
POTASSIUM DIHYDROGEN PHOSPHATE	186 I g
POTASSIUM HYDROGEN PHTHALATE	185 i

Certified By: Thermo Fisher Scientific

Water and Lab Products

Quality Assurance

Certificate Date: 08/17/23

244721-001 Rev.K

Water and Lab Products

Australia: (613) 9757-4300 In Australia: (1300) 735-295 China: (86) 21-6865-4588
Germany: (49) 6184-90-6321 India: (91) 22-6716-2261/2247 Japan: (81) 045-453-9175
North America: 1-978-232-6000 Toll Free: 1-800-225-1480 Singapore: (65) 6778-6876

Find out more at thermofisher.com/water